

Subject : Pre- Bid meeting minutes of selection for “Supply And Delivery Of Testing Equipment For Pedestal Fan Test Lab”

Date : 01st September 2026

Time : 10.30AM

Venue : DG Room, SLSEA

NO	Bidders request	Reply
1	Is it need to have ISO 17025 accredited calibration certificate Voltage Stabilizer and Variable transformer in table 02 and 07	Yes, ISO 17025 accredited calibration certificate needed for voltage stabilizer and variable transformer.
2	What is the power flow of system?	Power flow of the system as follows, Voltage stabilizer → Variable transformer ↓ Power analyzer
3	What is the instrument delivery period	Instrument delivery period 60 days
4	What is the delivery location	Instrument delivery place will be Sri Lanka Sustainable Energy Authority, No.80, Sri Ernest De Silva Mawatha, Colombo 07.
5	Is the compatible software for	The data acquisition unit shall be supplied with compatible software with a minimum validity period of two (2) years. The software shall also have an option for renewal after the initial two-year period.